

Technical Specifications	
Channel	Two, Each channel can be adjusted independently.
Output Mode	TENS, NMES
Output Waveform	Symmetrical biphasic asynchronous
Pulse Duration	Adjustable 50~400 μ s
Pulse Frequency	Adjustable 1Hz~120Hz
Output Intensity	Adjustable 0mA~100mA

Longest

Guangzhou Longest Science & Technology Co., Ltd.
 Address: 5&6F, Building B4, No.11, Kaiyuan Avenue, Science City,
 Guangzhou Hi-tech Industrial Development Zone, 510530 Guangzhou, China
 Tel: 86-20-66353999 Fax: 86-20-66353920
 E-mail: export@longest.cn Website: www.gzlongest.com



MStim Reha LGT-231

For Dysphagia Treatment




Health Life Longest Care

Longest

MStim Reha LGT-231

For Dysphagia Treatment


Mstim Reha LGT-231 utilizes the principle of NMES (Neuromuscular Electrical Stimulation), for prevention and retardation of disuse atrophy, for muscle re-education, and for relaxation of muscle spasms in the treatment of swallowing dysfunction in impaired neuromuscular function.



+ Applications

- ② Suitable for reconstruction of swallowing function after central nervous system injury or neck injury due to various reasons (cerebrovascular accident, brain trauma, etc.)

- NMES enhances the strength of the normal muscles associated with swallowing and prevent disuse atrophy.




- Assist swallowing training to avoid asphyxia and pneumonia caused by dysphagia, helping patients to swallow food safely.
- Form a normal and a good quality swallowing mode.

+ Features


- ② Ingenuity Design
Small, convenient for placing and carrying



- ② Intelligent Bluetooth Connection
Quick connection easy and convenient



- ② User-defined Programs
Widely adjustable parameter to meet different clinical needs



- ② Rich Preset Protocols
Colorful muscle patch diagram to facilitate the correct placement of muscle position, and referential clinical parameters



- ② Magnetic Electrodes
Easy and quick connection and removal of electrodes



- ② Real-time Display
Visualization of the training process

